

**DEPARTMENT OF ENVIRONMENTAL QUALITY - WATER DIVISION**  
**EQUIPMENT TEMPERATURE LOG/THERMOMETER VERIFICATION CHECK SHEET**  
10-06

FACILITY NAME:		VPDES NO:	<b>VA00</b>		DATE:								
EQUIPMENT	RANGE	IN RANGE		INSPECT READING °C	CHECK & LOG DAILY		CORRECT INCREMENT		ANNUAL THERMOMETER VERIFICATION				
									Is the NIST / NIST Traceable Reference Thermometer within the manufacturer's expiration date or recertified yearly?				Y/N
		Y	N		Y	N	Y	N	DATE CHECKED	MARKED Y    N		CORR FACTOR	INSPECT TEMP °C
SAMPLE REFRIGER.	1-4°C												
AUTO SAMPLER	1-4° C												
BOD INCUBATOR	20° C ± 1° C												
SOLIDS DRYING OVEN	103-105° C												
WATER BATH	44.5° ± .2° C												
INCUBATOR	35°± .5° C												
AUTOClave	121° C IN 30 MIN												
HOT AIR STERILIZING	170° ± 10° C												
O & G WATER BATH	70°± 2° C												
REAGENT REFRIGER.	1-4° C												
pH METER	± 1° C												
DO METER	± 1° C												
THERMOMETER-OUTFALL	± 1° C												
Hg WATER BATH	95 °C												

PROBLEMS:  
**templog**